

X-ray Machine Inspection Report

Registrant Name _____
 Address _____
 City, State, ZIP _____
 Phone (____) _____
 Surveyor Name _____
 Surveyor Signature _____
 Survey Date _____ Office Contact Person _____

Specific Type of Practice
 _____ Hospital
 _____ Medical
 _____ Dental
 _____ Veterinary
 _____ Podiatric
 _____ Chiropractic
 _____ Other (specify) _____

_____ Total # of X-ray tubes at this location
 _____ # of tubes inspected
 _____ # of serious violations discovered

GENERAL ADMINISTRATIVE ITEMS

OK N/A N/S S

- | | | | | |
|-------|-------|-------|-------|---|
| _____ | _____ | _____ | _____ | 1. Form RH-F-12, "Notice to Employees", not posted (12 VAC 5-481-2590C)
NON-SERIOUS |
| _____ | _____ | _____ | _____ | 2. Personnel radiation dose limits exceeded (explains in remarks) (12 VAC 5-481-640, 700 & 710)
SERIOUS |
| _____ | _____ | _____ | _____ | 3. Radiation exposure levels in unrestricted are excessive (explain in remarks) (12 VAC 5-481-720)
SERIOUS |
| _____ | _____ | _____ | _____ | 4. Personnel monitoring records inadequate (explain in remarks) (12 VAC 5-481-680F)
SERIOUS |
| _____ | _____ | _____ | _____ | 5. Personnel monitoring is necessary but is not properly provided (12 VAC 5-481-760)
SERIOUS |
| _____ | _____ | _____ | _____ | 6. Safety procedures not posted, inadequate, or not available (12 VAC 5-481-1590-A4)
NON-SERIOUS |
| _____ | _____ | _____ | _____ | 7. Gonadal shielding is necessary, but not available (12VAC 5-481-1590-A6) SERIOUS |
| _____ | _____ | _____ | _____ | 8. The speed of the film/screen used is not the fastest consistent with the diagnostic objective (12 VAC 5-481-1590 A9A) NON-SERIOUS |
| _____ | _____ | _____ | _____ | 9. Information and maintenance records inadequate or not available (12 VAC 5-481-1590 A12) NON-SERIOUS |
| _____ | _____ | _____ | _____ | 10. Registration form not available for review (12 VAC 5-481-370B) NON-SERIOUS |
| _____ | _____ | _____ | _____ | 11. X-ray log inadequate or not available (12 VAC 5-481-1590 A13) NON-SERIOUS |
| _____ | _____ | _____ | _____ | 12. Operator's list not posted, inadequate, or not available (12VAC 5-481-1590 A 14)
NON-SERIOUS |
| _____ | _____ | _____ | _____ | 13. Other / Remarks: _____ |

